


Thermo Scientific Orion 2002SS Stainless Steel Online 2-Electrode Conductivity Cell



General Information

The 2002SS 2-electrode conductivity cell is stainless steel, with a cell constant of 0.1 cm^{-1} . This sensor is intended for low conductivity or pure water samples in a process environment.

Calibration


Periodic calibration of the conductivity cell in a Thermo Scientific Orion conductivity standard is recommended. Choose a standard that is closest to the expected sample value. Calibration frequency depends on sample and measurement conditions. Refer to the appropriate monitor user guide for detailed calibration and wiring instructions.

Specifications

Body Material:	Stainless steel
Electrode Type:	2-electrode cell
Electrode Material:	Stainless steel
Measuring Range:	0.01 $\mu\text{S}/\text{cm}$ to 300 $\mu\text{S}/\text{cm}$
Nominal Cell Constant:	0.1 cm^{-1}
Temperature Range:	-5 to 100 °C
Automatic Temperature Compensation:	Yes, 30 k Ω NTC
Maximum Pressure:	100 psi (6.9 bar)
Cable Length:	5 meters (16.4 ft)

Cleaning

Conductivity cells should be cleaned with deionized water or with warm water mixed with liquid household detergent. If contaminants cannot be removed, soak the cell in 0.1 M HCl for 10 minutes, followed with a deionized water rinse. It is recommended that the cell be recalibrated after cleaning.



Legend
 — Conventional 2-Electrode Cells
 — DuraProbe 4-Electrode Cells
 — Sample Conductivity Range

Chart 1 – Typical Operating Ranges of Conductivity Cells

Ordering Information

Cat. No.	Description
2002SS	Stainless steel 2-electrode conductivity cell with 1 x 60 mL bottle of 100 μS/cm conductivity standard
2002CC	Carbon 4-electrode conductivity cell with 1 x 60 mL bottle each of 100 μS/cm, 1413 μS/cm 12.9 mS/cm and 111.9 mS/cm conductivity standard
2002BK	Bracket kit for pipe mounting
2002PK	Panel mounting kit
2002SH	Protective stainless steel housing
2002FC	Flowcell and adapters for 2002SS and 2002CC conductivity cells
011005	111.9 mS/cm conductivity standard, 5 x 60 mL bottles
011006	12.9 mS/cm conductivity standard, 5 x 60 mL bottles
011007	1413 μS/cm conductivity standard, 5 x 60 mL bottles
011008	100 μS/cm conductivity standard, 5 x 60 mL bottles

©2008 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries.



Environmental Instruments
 Water Analysis Instruments

North America
 166 Cummings Center
 Beverly, MA 01915 USA
 Toll Free: 1-800-225-1480
 Tel: 1-978-232-6000
 Dom. Fax: 1-978-232-6015
 Int'l Fax: 978-232-6031

Europe
 P.O. Box 254, 3860 AG Nijkerk
 Wallerstraat 125K, 3862 BN
 Nijkerk, Netherlands
 Tel: (31) 033-2463887
 Fax: (31) 033-2460832

Asia Pacific
 Blk 55, Ayer Rajah Crescent
 #04-16/24, Singapore 139949
 Tel: 65-6778-6876
 Fax: 65-6773-0836

255061-001 Rev.A 03-08

www.thermo.com/water

